

**Streptococcus pyogenes**  
**Detection with real time PCR reagents**

Overview:..... 1

Products..... 2

S. pyogenes FAM-BHQ1                      PP2800    0.055ml ..... 2

AttoMaster 2X Mix for qPCR              AM10      1.25 ml..... 2

S. pyogenes Plasmid 200 pg/ml            PLAS2800 0.25ml ..... 2

Detection of S.pyogenes DNA:..... 3

Thermal cycle conditions for PCR reactions on RotorGene\* ..... 3

Thermal cycle conditions for PCR reactions on LightCycler ..... 3

Use of the Streptococcus pyogenes plasmid:..... 3

Attostar Primers/probe PP2800 RotorGene-PCR 20µl reactions ..... 4

Attostar Primers/probe PP2800 LightCycler-PCR 20 µl reactions ..... 4

Extraction / amplification control with T4 bacteriophage ..... 5

Brief procedure for use of T4 as extraction and amplification control:..... 5


T4 multiplex reaction (PP160)..... 5

RotorGene Multiplex PCR 20µl reactions..... 5

Reference: ..... 6

Enlarged image: ..... 6

**Overview:**



Streptococcus pyogenes, Group A strep, causes the common tonsillitis, pharyngitis and impetigo. Group A strep is also the cause of serious, life-threatening, cellulitis and systemic disease. Immune processes are responsible for post-streptococcal rheumatic fever and glomerulonephritis.

Primers and probe are available for amplification and detection of Group A strep. The Group A strep primers are specific for a 98 bp portion of the SpecB gene. The Group A strep detection probe is a molecular beacon (1)

## Products

<b>Products</b>	<b>Catalog No.</b>	<b>Quantity</b>
<b>S. pyogenes FAM-BHQ1</b> <i>Store at -20C. 25- 20µl reactions</i> <i>Typical use: Add 2X master mix and DNA, then thermal cycle.</i> <i>Attostar reagent contains primers and probe.</i> <i>Detection at 510nm.</i>	<b>PP2800</b>	<b>0.055ml</b>
<b>AttoMaster 2X Mix for qPCR</b> <i>Store at -20°C. 125-20 ul reactions</i> <i>Contains Taq polymerase (requires heat activation), dNTPs (0.4 mM) with optimal dUTP to dTTP ratio, heat labile UDG, Mg(6 mM), and buffer.</i> <i>Typical use: Add Attostar Primer-probe, DNA, and then thermal cycle.</i>	<b>AM10</b>	<b>1.25 ml</b>
<b>S. pyogenes Plasmid 200 pg/ml</b> <i>Store at -20C.</i> <i>Typical use: make serial 10 fold dilutions in TE for standard curve, diluting 5 ul into 45 ul TE buffer.</i>	<b>PLAS2800</b>	<b>0.25ml</b>

## Detection of *S.pyogenes* DNA:

---

### Thermal cycle conditions for PCR reactions on RotorGene\*

\*Similar cycle conditions and reaction volumes may be used on many other thermal cyclers.

25°C 10 min (UDG treatment time)  
95°C 120 sec (activation for AttoMaster polymerase)  
40 cycles  
    95°C 15 seconds  
    60°C 30 seconds RotorGene Channel Setup FAM/Sybr, Cy5; Gain 7  
    72°C 30 seconds

FAM/Sybr has a source of 470nm and Detector 510nm (LightCycler use F1)

Cy5 has a source of 625nm and Detector 660hp nm

Quasar 670 has the same fluorescent absorption and emission as Cy5.

---

### Thermal cycle conditions for PCR reactions on LightCycler

25°C 10 min (UDG treatment time)  
95°C 120 sec (activation for AttoMaster polymerase)  
40 cycles  
    95°C 15 seconds  
    60°C 30 seconds acquire fluorescent signal on F1 gain =1  
    72°C 30 seconds  
40°C 30 seconds cool

---

### Use of the *Streptococcus pyogenes* plasmid:

Dilute the plasmid in TE to prepare a standard curve. Common dilutions would be 10-fold from 200 to 0.002pg/ml. The 0.02 pg/ml plasmid dilution contains 12 copies of plasmid in 2 µl.

.....

**Attostar Primers/probe PP2800 RotorGene-PCR 20µl reactions**

RotorGene Reagents needed for 20 ul PCR final reaction tube volumes										
Reaction tube number	1	2	3	4	5	6	7	8	9	10
Attostar Primer-Probe FAM labeled (10X)	2	4	6	8	10	12	14	16	18	20 µl
Master mix (2X)	10	20	30	40	50	60	70	80	90	100 µl
Dispense 12 ul / reaction tube										
Add 8 ul DNA / reaction tube										

**Attostar Primers/probe PP2800 LightCycler-PCR 20 µl reactions**

LightCycler Reagents needed for 20 ul PCR final reaction tube volumes										
Reaction tube number	1	2	3	4	5	6	7	8	9	10
Attostar Primer-Probe (10X)	2	4	6	8	10	12	14	16	18	20 µl
Master mix (2X)	10	20	30	40	50	60	70	80	90	100 µl
BSA 1 mg/ml	1	2	3	4	5	6	7	8	9	10 µl
Dispense 13 ul / reaction tube										
Add 7 ul DNA / reaction tube										

.....

**Extraction / amplification control with T4 bacteriophage**

Adding T4 bacteriophage (BAC130) to the sample provides DNA for extraction, amplification, and reaction condition PCR controls.

When added to a sample, T4 adds a known amount of DNA. The T4 DNA can then be extracted, amplified, and detected as a control. T4 controls for the efficiency of DNA extraction, the presence of PCR amplification inhibitors, intact amplification reagents ( DNA polymerase, buffer, dNTPs), and instrument function (thermal cycling and fluorescent detection system).

The T4 DNA may be detected in a separate PCR reaction (using FAM labeled T4 probe in PP100). Or the T4 DNA and test organism DNA may be detected using a multiplex reaction (using Quasar 670 labeled T4 probe in PP160 and FAM labeled test organism probe).

**Brief procedure for use of T4 as extraction and amplification control:**

- Add 5µl T4 bacteriophage to the sample. Proceed with DNA extraction. Dilutions of the bacteriophage may be made to give a final PCR Ct value that is about 35. At this dilution, the phage is more sensitive, i.e. more likely, to detect a poor extraction or the presence of PCR inhibitors in the reaction.

Please refer to the BAC130, PP100, and PP160 product literature.

**T4 multiplex reaction (PP160)**

**RotorGene Multiplex PCR 20µl reactions**

Multiplex Reagents needed for 20 ul PCR final reaction tube volumes										
Reaction tube number	1	2	3	4	5	6	7	8	9	10
Attostar Primer-Probe PP160	2	4	6	8	10	12	14	16	18	20 µl
Primer-Probe mix for TEST organism	2	4	6	8	10	12	14	16	18	20 µl
2 x master mix	10	20	30	40	50	60	70	80	90	100 µl
Dispense 14 ul / reaction tube										
Add 6 ul DNA / reaction tube										

Reaction tube number	11	12	13	14	15	16	17	18	19	20
Attostar Primer-Probe PP160	22	24	26	28	30	32	34	36	38	40 µl
Primer-Probe mix for TEST organism	22	24	26	28	30	32	34	36	38	40 µl
2 x master mix	110	120	130	140	150	160	170	180	190	200 µl
Dispense 14 ul / reaction tube										
Add 6 ul DNA / reaction tube										

.....

**Reference:**

- 1) <http://www.molecular-beacons.org/Introduction.html>
- 2) Image:S. pyogenes beta hemolytic colonies:  
<http://www.jnu.ac.in/Faculty/ajohri/>


---

For research use only. Not intended for any animal or human therapeutic or diagnostic use. This product is sold under license from the Public Health Research Institute. It may be used under PHRI Patent Rights only for the purchaser's research and development activities. 'Black Hole Quencher,' 'CAL Fluor,' 'Pulsar' and 'Quasar' are trademarks of and licensed by Biosearch Technologies, Inc., Novato, CA. The BHQ, CAL Fluor, Pulsar and Quasar dye technology is the subject of existing or pending patents including US Patent No. 7,019,129 and is licensed and sold under agreement with Biosearch Technologies, Inc.

For technical support, contact [Attostar@Attostar.com](mailto:Attostar@Attostar.com)

**Enlarged image:**

Beta hemolytic colonies of *S. pyogenes* on sheep blood plate (2).



10-18-07